



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	TANGEDCO TESTING LABORATORY, 129, WALLAJAH ROAD, CHENNAI, TAMIL NADU, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10971	Page No	1 of 5
Validity	06/09/2022 to 05/09/2024	Last Amended on	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Limits of Error and Interpretation of Test Results	IS 16444 (Clause 6.12)
2	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Limits of Error	IS 16444 (Clause 6.12)
3	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Meter Constant	IS 16444 (Clause 6.12)
4	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	No Load Test	IS 16444 (Clause 6.12)
5	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Repeatability of Error	IS 16444 (Clause 6.12)
6	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Starting Condition	IS 16444 (Clause 6.12)
7	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Test of Influence of Supply Voltage	IS 16444 (Clause 6.10.2)
8	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Test of Influence Quantities - 10% of 3rd Harmonics in Current	IS 16444 (Clause 6.12)
9	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Test of Influence Quantities - Frequency Variation	IS 16444 (Clause 6.12)
10	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Test of Influence Quantities - Reversed Phase Sequence	IS 16444 (Clause 6.12)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	TANGEDCO TESTING LABORATORY, 129, WALLAJAH ROAD, CHENNAI, TAMIL NADU, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10971	Page No	2 of 5
Validity	06/09/2022 to 05/09/2024	Last Amended on	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
11	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Test of Influence Quantities - Voltage Unbalance	IS 16444 (Clause 6.12)
12	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Direct Connected Watthour Smart Meter Class 1 and 2	Test of Influence Quantities - Voltage Variation	IS 16444 (Clause 6.12)
13	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Limits of Error	IS 16444 (Part 2, Clause 6.12)
14	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Limits of error and Interpretation of Test Results	IS 16444 (Part 2, Clause 6.12)
15	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Meter Constant	IS 16444 (Part 2, Clause 6.12)
16	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	No Load Test	IS 16444 (Part 2, Clause 6.12)
17	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Repeatability of Error	IS 16444 (Part 2, Clause 6.12)
18	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Starting Condition	IS 16444 (Part 2, Clause 6.12)
19	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Test of Influence of Supply Voltage	IS 16444 (Part 2, Clause 6.10.2)
20	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Test of Influence Quantities - 10% of 3rd Harmonics in Current	IS 16444 (Part 2, Clause 6.12)
21	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Test of Influence Quantities - Frequency Variation	IS 16444 (Part 2, Clause 6.12)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	TANGEDCO TESTING LABORATORY, 129, WALLAJAH ROAD, CHENNAI, TAMIL NADU, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10971	Page No	3 of 5
Validity	06/09/2022 to 05/09/2024	Last Amended on	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
22	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Test of Influence Quantities - Reversed Phase Sequence	IS 16444 (Part 2, Clause 6.12)
23	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Test of Influence Quantities - Voltage Unbalance	IS 16444 (Part 2, Clause 6.12)
24	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	A.C. Static Transformer Operated Watthour and Var-Hour Smart Meters, Class 0.2s, 0.5s and 1.0s	Test of Influence Quantities - Voltage Variation	IS 16444 (Part 2, Clause 6.12)
25	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Limits of Error	IS 13779 (Clause 11.1)
26	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Limits of Error and Interpretation of Test Results	IS 13779 (Clause 12.16)
27	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Meter Constant	IS 13779 (Clause 12.15)
28	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	No Load Test	IS 13779 (Clause 12.13)
29	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Repeatability of Error	IS 13779 (Clause 12.17)
30	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Starting Condition	IS 13779 (Clause 12.14)
31	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Test of Influence of Supply Voltage	IS 13779 (Clause 12.7.2)
32	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Test of Influence Quantities - 10% of 3rd Harmonics in Current	IS 13779 (Clause 11.1, 11.2 & 12.11)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	TANGEDCO TESTING LABORATORY, 129, WALLAJAH ROAD, CHENNAI, TAMIL NADU, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10971	Page No	4 of 5
Validity	06/09/2022 to 05/09/2024	Last Amended on	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
33	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Test of Influence Quantities - Frequency Variation	IS 13779 (Clause 11.1, 11.2 & 12.11)
34	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Test of Influence Quantities - Reversed Phase Sequence	IS 13779 (Clause 11.1, 11.2 & 12.11)
35	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Test of Influence Quantities - Voltage Unbalance	IS 13779 (Clause 11.1, 11.2 & 12.11)
36	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter of Class 1 and 2	Test of Influence Quantities - Voltage Variation	IS 13779 (Clause 11.1, 11.2 & 12.11)
37	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	15-Minute Block Average Frequency Measurement Test	CBIP No.:325 (Clause 6.9 (c)(i))
38	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Limits of Error	IS 14697 (Clause 11.1)
39	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Limits of Error and Interpretation of Test results	IS 14697 (Clause 12.15)
40	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Meter Constant	IS 14697 (Clause 12.14)
41	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	No Load Test	IS 14697 (Clause 12.12)
42	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Reactive Energy Registration - Verification Test during Voltage High (V>103% and Voltage Low (V	CBIP No.:325 (Clause 6.1 (c) (iv & v))
43	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Repeatability of Error	IS 14697 (Clause 12.16)



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	TANGEDCO TESTING LABORATORY, 129, WALLAJAH ROAD, CHENNAI, TAMIL NADU, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10971	Page No	5 of 5
Validity	06/09/2022 to 05/09/2024	Last Amended on	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
44	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Starting Condition	IS 14697 (Clause 12.13)
45	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Test of Influence of Supply Voltage	IS 14697 (Clause 12.7.2)
46	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Test of Influence Quantities - 10% of 3rd Harmonics in Current	IS 14697 (Clause 11.1, 11.2 & 12.10)
47	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Test of Influence Quantities - Frequency Variation	IS 14697 (Clause 11.1, 11.2 & 12.10)
48	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Test of Influence Quantities - Reversed Phase Sequence	IS 14697 (Clause 11.1, 11.2 & 12.10)
49	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Test of Influence Quantities - Voltage Unbalance	IS 14697 (Clause 11.1, 11.2 & 12.10)
50	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Transformer Operated Energy Meter of Class 0.2S, 0.5S and 1S	Test of Influence Quantities - Voltage Variation	IS 14697 (Clause 11.1, 11.2 & 12.10)